PROCESSED AT :

Thyrocare
Tests you can trust

Thyrocare Technologies Limited, D-37/3, TTC MIDC, Turbhe, Navi Mumbai - 400 703 9870666333 wellness@thyrocare.com

First National Diagnostic Chain to have 100% of its Labs with NABL Accreditation#

NAME : XXXXXXXXXXXXXXXXXXXX
REF. BY : XXXXXXXXXXXXXXXXXXXX
TEST ASKED : HbA1c with Graph

SAMPLE COLLECTED AT :
XXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXXX

TEST NAME	TECHNOLOGY	VALUE	UNITS
HbA1c - (HPLC)			
	H.P.L.C	< 4	%

Bio. Ref. Interval. :

Bio. Ref. Interval.: As per ADA Guidelines
Below 5.7% : Normal
5.7% - 6.4% : Prediabetic
>=6.5% : Diabetic

Guidance For Known Diabetics
Below 6.5% : Good Control
6.5% - 7% : Fair Control
7.0% - 8% : Unsatisfactory Control
>8% : Poor Control

Method : Fully Automated H.P.L.C method

AVERAGE BLOOD GLUCOSE (ABG)	**CALCULATED**	**68**	**mg/dL**


Bio. Ref. Interval. :

90 - 120 mg/dl : Good Control
121 - 150 mg/dl : Fair Control
151 - 180 mg/dl : Unsatisfactory Control
> 180 mg/dl : Poor Control

Method : Derived from HBA1c values

Please correlate with clinical conditions.

ALERT !!!VALUES TOO LOW. HEMOGLOBIN ELECTROPHORESIS SUGGESTED FOR RULING OUT A HEMOGLOBIN ABNORMALITY. ALERT !!!VALUES TOO LOW. HEMOGLOBIN ELECTROPHORESIS SUGGESTED FOR RULING OUT A HEMOGLOBIN ABNORMALITY. ALERT !!!VALUES. HEMOGLOBIN ELECTROPHORESIS.



Peak Name	RT (sec)	Area	Area%	Concentration (% NGSP)
A1a	5.69	1333.52	1.20	---
A1b	7.20	1103.49	0.99	---
F	8.24	936.06	0.84	---
LA1c	11.44	1801.38	1.62	---
HbA1c	13.78	5175.46	---	6.0
P3	18.72	5910.14	5.30	---
A0	24.39	95260.12	85.42	---
Total Area:		111520		

Dummy Report

~~ End of report ~~

Sample Collected on (SCT) : Sample collection time
Sample Received on (SRT) : Sample receiving time at Lab
Report Released on (RRT) : Report release time
Sample Type : EDTA Whole Blood
Labcode :
Barcode :

Doctor 1 Sign Doctor 2 Sign

Scan QR code to verify authenticity of reported results; active for 30 days from release time.